

# Calibration Results Table

**Photocopy this page!**

Photocopy this page before you begin your calibration. That way, you'll have fresh copies of this results table handy whenever you need them. Record your calibration results on one of the photocopies as you work your way through the calibration process.

---

Analyzer Model Signature 5000

Calibration Date: \_\_\_/\_\_\_/\_\_\_

Technician: \_\_\_\_\_

Tester Serial Number: \_\_\_\_\_

Calibration Step	Result Seen	Correct Result	Pass/Fail
No Connections		NO CONNECTS	
Begin at J1-J2		2DAE40	
Begin at J3-J4		89874E	
Begin at J5-J6		1852F0	
Begin at J7-J8		677C81	
Complete at J1-J2		86FD14	
Complete at J3-J4		011394	
Complete at J5-J6		FCCC8F	
Complete at J7-J8		85EDF2	